Shared Equipment Cleaning

****AFTER EVERY PATIENT USE****

Shared Equipment	Nurses/Point of Care Staff		EVS
	Cavi Wipes (use with gloves)	**Alcohol Wipes** (70% Isopropyl Alcohol)	Terminal Cleaning and MDRD
Glucometers	√	√	
• avoid liquid entering electrical connections & ports • weekly cleaning of docking stations and carrying totes required			
Vital Signs Machines	√	√	
Bladder Scanners	*	√	
• clean the device and the probe, do not immerse or allow liquids to leak into device			
ECG machine	*	√	
• clean the cables and interface module surface, do not wet the connectors			
Oxygen Saturation Monitor	√ (pack, wires)	√ (probe)	
Thermometers	√		
Gym Equipment	√		
Mechanical Lifts	√		
Transferring Equipment	√		
IV pump	√		√ (place in soiled utility)
Kangaroo pump	√		√ (place in soiled utility)
CADD pump	√		
Commodes	√		Terminal Cleaned in rooms after discharge
Wheelchairs	√ (touchpoints)		
• external contractor cleans wheelchairs after patient discharge with wheelchair cleaning machine			
Med Carts	√ (surface) √ (cassette drawers)		Yearly deep clean
Isolation Carts			√
AirMattress Pumps			√
AirMattress Covers			√ (washed through internal laundry dept.)
Weighing Scales	√		
ASCOM Nurse Call Phone	√		
Laundry/Waste hampers	√		
Supply Carts/linen	√		
Stethoscopes	√		
Personal Cell Phones	Please check with your manufacturer – this is a reminder that personal cell phones require regular cleaning if you are using them during work hours		

Pumps (IV, Kangaroo) are to be brought to soiled utility rooms after patient discharge – cleaned by Medical Device Reprocessing Department (MDRD). Clean pumps can be retrieved from CMED.

CaviWipes – contain: Diisobutylphenoxyethoxyethyl dimethyl benzyl ammonium chloride and isopropanol

*considered too abrasive for the screen portion of this equipment (manufacturer guidelines) but CaviWipes can be used if necessary